

Notice of Allowability

Application No.

10/620,057

Examiner

Emily Y. Chan

Applicant(s)

THURAIRAJARATNAM ET AL.

Art Unit

2829

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 3-22-05.
2. ☒ The allowed claim(s) is/are 1,5-10,15 and 19-23.
3. ☒ The drawings filed on 5-6-04 are accepted by the Examiner.
4. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).
 - * Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

5. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
6. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.

Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
7. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☒ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☒ Information Disclosure Statements (PTO-1449 or PTO/SB/08),
Paper No./Mail Date 04/04
4. ☐ Examiner's Comment Regarding Requirement for Deposit
of Biological Material
5. ☐ Notice of Informal Patent Application (PTO-152)
6. ☐ Interview Summary (PTO-413),
Paper No./Mail Date _____
7. ☐ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____

DETAILED ACTION

Allowable Subject Matter

1. Claims 1,5-10,15 and 19-23 are allowed over prior art.

The following is a statement of reasons for the indication of allowable subject matter:

2. The reasons for allowance of the claims are due to the amended independent claims 1 and 15 with specific structure connection of the Digital Sampling Oscilloscope (DSO) recited in the claims. The prior art of the record does not teach or suggest that the DSO is connected to both tester and test head so that the DSO can launch a signal to the test head and receive a reflected signal from the test head and can provide the signals to the tester. The persuasive arguments are found on the "Remarks" section on page 7. Claims 5-10 and 19-23 are dependent on claims 1 and 15 respectively and are allowed accordingly.
3. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

Swart US Patent No. 6,191,601 discloses test fixture matched impedance testing comprising a TDR meter (46) connected to a tester (20).

Conclusion


Any inquiry concerning this communication or earlier communications from the examiner should be directed to Emily Y. Chan whose telephone number is 571-272-1956. The examiner can normally be reached on 8:30-5:30.

Art Unit: 2829

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nestor Ramirez can be reached on 571-272-2034. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

EC
5-14-05


VINH NGUYEN
PRIMARY EXAMINER
A.U. 2829
05/16/05